



Quarterly Report for Bazan Fence line Monitoring System – Quarter 4 2022





January 06, 2023
Bazan Group
Haifa, Israel

Subject: Quarterly Report for Open-Path UV Air Monitoring

This report summarizes the data collected by the four open-path UV air monitoring systems that were installed at the Bazan refinery during the period of October to December 2022 (Quarter 4). During this time, the four systems continuously collected and quantified the concentration of Benzene, Toluene, Ethyl Benzene and Xylene (BTEX) gas in the ambient air. Information from the analyzers was reported along with meteorological data to a secured website, and alarm notifications were generated in the event target gases were detected above preset levels. Summary information is included for each beam path. The following report presents the summary results of the measurement period.

Operational Performance Events

As of the end of Q4 2022 the onstream efficiency for the systems was 97.81%.

Maintenance Activities

Routine maintenance and quality assurance/quality control (QA/QC) for the open path UV monitoring systems occurred on 23 October, 25 October, 27 October, 30 October, 31 October, 14 November, 22 November, 24 November, 27 November, 30 November, 16 December, 21 December, 29 December and 30 December 2022. Scheduled maintenance for lamp replacements took place on 18 October 2022. On site annual maintenance took place on 8 November, 9 November, 10 November, 13 November, 14 November, 15 November and 22 November 2022.



Summary Findings

From the results of the report the following were noted:

- The sample paths detected compounds at different times. This was expected as the paths (due to their orientation) were affected by sources from the refinery under different weather conditions.
- Winds were predominantly from the South-South-East during the period under review.

Please do not hesitate to contact me if you have questions or need additional information regarding the report.

Best Regards,
Bazan Group



Report Details

Reference	Bazan_Quarterly report_Q4_2022
Report Title	Quarterly Report for Bazan Fence line Monitoring System – Quarter 4 2022
Date Submitted	January 06, 2023
Client	Israel Ministry of Environmental Protection
Prepared by	Mark Baird Email: mwbaird@argos-sci.co.za Signed:
Review by	Amir Sagi Email: samir@bazan.co.il Signed:
Approval	Eliko Angel Email: aeliko@bazan.co.il Signed:
Status	Signed
ISRAC ISO/IEC 17025 accreditation	<p>The use of ISRAC symbol relates to tests/calibrations which are included in the organization scope of accreditation, and performed according to the accreditation rules as detailed in the accreditation certificate. <u>The scope of accreditation is for Benzene results only</u>, any other information is not part of the accreditation.</p> <p>ISRAC is not responsible for the testing results conducted by the CAB and the CAB's accreditation is not considered as an approval of ISRAC or a different party related to the assessed. ISRAC accreditation is not considered as an approval of either the CAB's procedures or its personnel.</p> <p>This document must be referred to entirely and copying of any part of it to other documents is forbidden.</p>
Statement concerning the	"The Israel Laboratory accreditation authority (ISRAC) is one of the signatories of the International Laboratory Accreditation Cooperation (ILAC) arrangement for the mutual recognition of Testing/Sampling/ Calibration/ Inspection results.



International Recognition	
Notice	<p>Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders September be prosecuted to the fullest extent of the law.</p>



Table of Contents

REPORT DETAILS	3
SECTION 1 – INTRODUCTION	6
SECTION 2 - QUALITY ASSURANCE QUALITY CONTROL CHECKS	8
SECTION 3 - SUMMARY OF FIELD DATA.....	15
UV1 – SUMMARY OF DAILY AVERAGE FIELD DATA	16
UV2 – SUMMARY OF REALTIME FIELD DATA	20
UV4 – SUMMARY OF REALTIME FIELD DATA	24
UV5 – SUMMARY OF REALTIME FIELD DATA	28
SECTION 4 – SUMMARY NOTES	33



Section 1 – Introduction

The purpose of the Argos open-path UV air monitoring project at the Bazan Refinery in Haifa, Israel is to measure Benzene, Toluene, Ethylbenzene and Xylene (BTEX) gases on a real-time basis and to present this information via a secured website. The system also send alerts for detection levels defined by Israeli authorities. The fence-line monitoring equipment is installed along four portions of the fence line at the Bazan refinery, these locations are shown in Figure 1.1. The beam paths covered are referenced as: Refinery Northeastern line 867 meters (UV_1); Gadiv Eastern line 428 meters (UV_2); Western Line (UV_5) 311 m and Refinery Southwestern line (UV_4) 822 meters.

Figure 1.1 - Location Map Showing Fence Lines Covered by Argos Open-path UV Air Monitoring Systems





Each site is equipped with a meteorological station measuring the following parameters:

- Wind speed
- Wind direction
- Temperature
- Relative humidity

The purpose of meteorological station is to help locate pollution source direction only on the Fenceline system height.

The quarterly summary of the measurements performed by the system for Q4 2022, are presented in the sections that follow. The measurements for Benzene were performed using an in-house validated method FLM-QLT-MET-001 Determination of BTEX by UV DOAS based on TO-16 and EPA 301 methods.

During November the system had maintenance audit under which the UV and the Met systems were checked and calibrated, these resulted better performance of the systems.



Section 2 - Quality Assurance Quality Control Checks

The Argos Open-path UV air monitoring systems employ a number of methods to check the data quality of the system. Table 2.1 summarizes the routine data quality checks employed during the project. Each data quality check is described below.

Table 2.1 –Quality Checks

Data Quality Check	Frequency
Continuous Data Quality Checks	Continuous
System Check	Daily
Challenge of System with Known Quantity of Gas	Every two weeks
Independent Check of Gas Detects	As Needed

Continuous Data Quality Checks (Frequency – Continuous)

Data generated by the fence-line monitoring equipment undergoes review throughout the measurement and reporting process. This includes automated QA/QC checks that occur before data is reported on the real-time website. Automated data checks are listed in Table 2.2

System Check (Frequency Daily)

During the measurement period Argos provided continual on-call support for the fence-line monitoring network. This includes an alarm system that notifies a support team in the event of instrument malfunction or high detections of gases, loss of Internet connections and other issues that might impact on the performance of the monitoring equipment. The support team includes staff scientists who are experts in the field for UV spectroscopy. In addition, each day Argos staff remotely access the local instrument computers and perform data checks to ensure the system is operating properly. This includes but was not limited to:

- Troubleshooting software issues
- Checking light signals
- Perform validation checks on gas detections



Table 2.2 – Real-time Data Quality Checks

Real-Time Check	Check	Action
Low Signal Alarm	Signal threshold test	If signal is below threshold value: 1) Real-time website reports "Low Signal" to analyzer. Automated email is sent to notify support staff of the issue.
Instrument Error Code	Instrument Error Code	Real-time website reports "off-line" message. Automated email is sent to notify support staff of the issue.
Instrument Workstation Off-line	Instrument Communication Check	Real-time website reports "off-line" message. Automated email is sent to notify support staff of the issue.
Internet Connection Lost	Backup Connection enabled	Automated email is sent to notify support staff of the issue.
High Detection	Valid Data Detection Above Threshold	Real-time website indicates detection above alarm threshold. Automated email is sent to notify support staff of the issue.

Challenge of System with Known Quantity of Gas (Frequency – Every two weeks)

The UV systems are calibrated by inserting a known concentration of a target gas into the beam and then measuring the system response. The target gas is held inside a sealed cell with windows that minimize absorption of UV light. The measure will be considered passing if the quantified result is within 15% of the expected value.



Tables 2.3, 2.4, 2.5 and 2.6 below summarises the QA check results for the system for Q4 2022.

Table 2.3: Summary of Q4 2022 Benzene Challenge Gas Checks at UV1

Date	Times	Expected (ppb)	Measured (ppb)	% Difference	Status
25 October 2022	16:13	76.24	67.39	11.6	Passed
25 October 2022	16:18	7.14	7.19	0.7	Passed
25 October 2022	16:23	3.57	3.43	3.9	Passed
31 October 2022	19:01	76.24	81.05	6.3	Passed
31 October 2022	19:07	7.14	7.35	3.0	Passed
31 October 2022	19:12	3.57	3.60	0.8	Passed
14 November 2022	16:05	76.24	76.02	0.29	Passed
14 November 2022	16:10	15.25	16.70	9.55	Passed
14 November 2022	16:16	7.62	6.38	16.28	Failed ¹
24 November 2022	15:25	76.24	75.98	0.34	Passed
24 November 2022	15:52	15.25	16.95	11.16	Passed
24 November 2022	15:59	7.62	6.19	18.86	Failed ¹
21 December 2022	11:52	76.24	76.35	0.1	Passed
21 December 2022	11:57	7.40	7.51	1.5	Passed
21 December 2022	12:02	4.38	5.03	14.5	Passed
29 December 2022	20:38	76.24	76.35	0.1	Passed
29 December 2022	20:43	15.25	13.65	10.5	Passed
29 December 2022	20:48	3.35	3.78	13	Passed

¹ Addressed as part of NCR 10_30_22



Table 2.4: Summary of Q4 2022 Benzene Challenge Gas Checks at UV2

Date	Times	Expected (ppb)	Measured (ppb)	% Difference	Status
23 October 2022	23:18	77.80	72.50	6.8	Passed
23 October 2022	23:23	4.35	4.81	10.8	Passed
23 October 2022	23:28	2.53	2.66	5.1	Passed
27 October 2022	09:24	77.80	67.92	12.7	Passed
27 October 2022	09:43	4.35	4.94	13.7	Passed
27 October 2022	09:36	2.53	2.51	0.7	Passed
27 November 2022	16:24	77.80	78.68	1.13	Passed
27 November 2022	16:29	15.56	10.19	34.50	Failed ²
27 November 2022	16:35	7.78	4.74	39.05	Failed ²
30 November 2022	08:20	77.80	78.49	0.89	Passed
30 November 2022	08:25	15.56	11.75	24.49	Failed ²
30 November 2022	08:30	7.78	6.28	19.25	Failed ²
16 December 2022	18:16	77.80	78.49	0.9	Passed
16 December 2022	18:21	9.97	10.11	1.0	Passed
16 December 2022	18:26	5.27	5.63	7.0	Passed
29 December 2022	15:48	77.80	67.96	12.7	Passed
29 December 2022	15:54	15.56	14.32	8.0	Passed
29 December 2022	15:59	3.81	4.35	14.2	Passed

² No blank cell inserted into the beam to collect a background from addressed as part of NCR 10_30_22



Table 2.5: Summary of Q4 2022 Benzene Challenge Gas Checks at UV4

Date	Times	Expected (ppb)	Measured (ppb)	% Difference	Status
25 October 2022	17:47	72.51	66.93	7.7	Passed
25 October 2022	17:52	3.81	3.28	14.1	Passed
25 October 2022	17:57	1.28	1.18	7.6	Passed
31 October 2022	14:12	72.51	61.66	15.0	Passed
31 October 2022	14:17	2.11	2.35	11.1	Passed
31 October 2022	14:22	1.28	1.28	0.3	Passed
22 November 2022	14:52	72.51	72.33	0.24	Passed
22 November 2022	14:58	14.50	15.72	8.43	Passed
22 November 2022	15:05	7.25	6.20	14.42	Passed
24 November 2022	15:21	72.51	72.34	0.23	Passed
24 November 2022	15:28	14.50	15.64	7.89	Passed
24 November 2022	15:38	7.25	6.27	13.54	Passed
16 December 2022	19:07	72.51	72.61	0.1	Passed
16 December 2022	19:12	14.50	12.96	10.6	Passed
16 December 2022	19:17	3.81	4.37	15.0	Passed
30 December 2022	12:58	72.51	72.61	0.1	Passed
30 December 2022	13:03	14.50	12.35	14.8	Passed
30 December 2022	13:08	7.25	6.22	14.0	Passed



Table 2.6: Summary of Q4 2022 Benzene Challenge Gas Checks at UV5

Date	Times	Expected (ppb)	Measured (ppb)	% Difference	Status
30 October 2022	14:07	100.32	88.48	11.8	Passed
30 October 2022	14:13	4.14	5.10	23.2	Failed ³
30 October 2022	14:36	2.18	2.68	22.8	Failed ³
31 October 2022	10:24	100.32	85.53	14.7	Passed
31 October 2022	10:29	8.73	10.68	22.3	Failed ³
31 October 2022	10:34	4.00	5.30	32.6	Failed ³
22 November 2022	17:19	100.32	100.43	0.11	Passed
22 November 2022	17:24	20.06	18.77	6.44	Passed
22 November 2022	17:30	10.03	11.21	11.79	Passed
30 November 2022	09:39	100.32	100.39	0.07	Passed
30 November 2022	09:44	20.06	19.38	3.40	Passed
30 November 2022	09:49	10.03	10.65	6.14	Passed
21 December 2022	13:03	100.32	95.4	4.9	Passed
21 December 2022	13:08	3.73	4.28	14.8	Passed
21 December 2022	13:13	1.52	1.56	3.0	Passed
29 December 2022	16:28	100.32	86.12	14.15	Passed
29 December 2022	16:33	12.54	12.31	2.0	Passed
29 December 2022	16:39	5.38	5.3	2.0	Passed

³ NCR 10_30_22 was raised in order to resolve this issue



Independent Check of Gas Detects

In addition to automated features in the software, Argos technical data analysts have the ability to check the system performance by carrying out independent quantification of target gases. Depending on the specific application, these activities are performed on a routine basis to ensure the automated data collection and verification process is functioning correctly. An example of this process is presented below:

- Collect a data spectrum in the atmosphere when the target gas is not present. Define this as the background spectrum.
- Collect a data spectrum in the atmosphere when the target gas is present. Define this as the data spectrum.
- Subtract the logarithms of the two spectra. This resulting spectrum is defined as an absorbance spectrum.
- Compare this spectrum to a quantitative absorbance spectrum of the target gas.

Minimum Detection Limit Checks

The minimum detection limits for the system were calculated for Q4 2022 and are shown in Table 2.7 below:

Table 2.7: Minimum Detection Limits for Systems for Q4 2022

Location	October MDL (ppb)	November MDL (ppb)	December MDL (ppb)
UV_1	0.67	0.38	0.67
UV_2	0.94	0.36	0.39
UV_4	0.57	0.75	0.70
UV_5	0.50	0.51	0.38



Section 3 - Summary of Field Data

As mentioned in Section 1, the air monitoring equipment operated continuously during Q4 2022. The following figures and tables summarize the data collected during this time period for BTEX gases. Each system collected data at five-minute averages. Table 3.1 lists the time periods that valid data was collected at each location.

Table 3.1 – Data Collection periods

System	Data Start	Data End	Onstream (%)
UV1	10/01/2022	12/31/2022	96.89
UV2	10/01/2022	12/31/2022	96.18
UV4	10/01/2022	12/31/2022	97.64
UV5	10/01/2022	12/31/2022	98.72

The following alarms were set by the client and were used to measure the performance of the system:

- Benzene:
 - 20 microgram/m³ half hour average. (30 min alarm)
 - 2 consecutive measurements of 10 microgram/m³ half hour average. (1 hr alarm)
 - 3.9 microgram/m³ daily average. (24 hr alarm)
- Toluene: 3,770 microgram/m³ daily average.
- Total Xylenes: 4,800 microgram/m³ daily average (Almog value).
- Ethyl Benzene: 54,000 microgram/m³ 15-minute average (Almog value)



UV1 – Summary of Daily Average Field Data

Figures 3.1 to 3.5 show the daily average data collected from the UV1 system

Figure 3.1 – Benzene 24 Hour Average Data for UV1 for Q4 2022






Figure 3.2 – Toluene 24 Hour Average Data for UV1 for Q4 2022




Figure 3.3 – Xylene 24 Hour Average Data for UV1 for Q4 2022






Figure 3.4 - Ethyl Benzene 24 Hour Average Data for UV1 for Q4 2022







Figure 3.5 – Wind Speed and Wind Direction for UV1 for Q4 2022





UV2 – Summary of Realtime Field Data

Figures 3.6 to 3.10 show the daily average data collected from the UV2 system

Figure 3.6 – Benzene 24 Hour Average Data for UV2 for Q4 2022






Figure 3.7 – Toluene 24 Hour Average Data for UV2 for Q4 2022




Figure 3.8 – Xylene 24 Hour Average Data for UV2 for Q4 2022






Figure 3.9 – Ethyl Benzene 24 Hour Average Data for UV2 for Q4 2022







Figure 3.10 – Wind Speed and Wind Direction for UV2 for Q4 2022





UV4 – Summary of Realtime Field Data

Figures 3.11 to 3.15 show the daily average data collected from the UV4 system

Figure 3.11 – Benzene 24 Hour Average Data for UV4 for Q4 2022






Figure 3.12 – Toluene 24 Hour Average Data for UV4 for Q4 2022




Figure 3.13 – Xylene 24 Hour Average Data for UV4 for Q4 2022




Figure 3.14 – Ethyl Benzene 24 Hour Average Data for UV4 for Q4 2022







Figure 3.15 – Wind Speed and Wind Direction for UV4 for Q4 2022





UV5 – Summary of Realtime Field Data

Figures 3.16 to 3.20 show daily average data collected from the UV5 system

Figure 3.16 – Benzene 24 Hour Average Data for UV5 for Q4 2022






Figure 3.17 – Toluene 24 Hour Average Data for UV5 for Q4 2022




Figure 3.18 – Xylene 24 Hour Average Data for UV5 for Q4 2022




Figure 3.19 – Ethyl Benzene 24 Hour Average Data for UV5 for Q4 2022







Figure 3.20 – Wind Speed and Wind Direction Data for UV5 for Q4 2022





The 24-hour alarm level was not exceeded for benzene during Quarter 4 2022.



Section 4 – Summary Notes

From the results in Section 3 above the following were noted:

- The sample paths detected compounds at different times. This was expected as the paths (due to their orientation) were affected by sources from the refinery under different weather conditions.
- Winds were predominantly from the South-South-East during the period under review.

The results of the measurements indicate the Argos Open-path UV air monitoring systems were able to detect and quantify BTEX emissions from sources both within and outside of the Bazan Refinery. Specific activities associated with detections by the fence-line systems could be identified based on the specific gases being detected, the wind speed and direction, and the specific time of day when the detection occurred.



Appendix A: Calibration Certificates



Sealed Calibration Cell

Certificate of Calibration

This calibration cell was individually calibrated by spectroscopic analysis using ultra-violet reference libraries created by permeation tubes traceable to N.I.S.T standards.

Permeation Tubes Serial Numbers: 49302/49303

Permeation Fluid: Benzene

Calibrated with 40.0 ppm-m Reference Standard

Calibration Cell Serial # **BENZ052522-05**

Calibration Cell Concentration: **66.1 ppm-m**

Lower 95% confidence: 64.1 ppm-m

Correlation Coefficient: 0.991

Upper 95% confidence: 68.1 ppm-m

Laboratory Manager: 

Date: May 25, 2022

Brentley S. Olive MSPH, Ph.D., CIH



DEPARTMENT of CHEMISTRY and INDUSTRIAL HYGIENE

College of Arts and Sciences

UNA Box 5049, Florence, AL 35632-0001

P: 256.765.4622 | F: 256.765.4958 | www.una.edu/chem-ih

Equal Opportunity / Equal Access Institution



Sealed Calibration Cell

Certificate of Calibration

This calibration cell was individually calibrated by spectroscopic analysis using ultra-violet reference libraries created by permeation tubes traceable to N.I.S.T standards.

Permeation Tubes Serial Numbers: 49302/49303

Permeation Fluid: Benzene

Calibrated with 20.0 ppm-m Reference Standard

Calibration Cell Serial # **BENZ072822-01**

Calibration Cell Concentration: **33.3 ppm-m**

Lower 95% confidence: 29.9 ppm-m

Correlation Coefficient: 0.970

Upper 95% confidence: 36.6 ppm-m

Laboratory Manager: _____

Date: July 28, 2022

Brentley S. Olive MSPH, Ph.D., CIH



DEPARTMENT of CHEMISTRY and INDUSTRIAL HYGIENE

College of Arts and Sciences

UNA Box 5049, Florence, AL 35632-0001

P: 256.765.4622 | F: 256.765.4958 | www.una.edu/chem-ih

Equal Opportunity / Equal Access Institution



Sealed Calibration Cell

Certificate of Calibration

This calibration cell was individually calibrated by spectroscopic analysis using ultra-violet reference libraries created by permeation tubes traceable to N.I.S.T standards.

Permeation Tubes Serial Numbers: 49302/49303

Permeation Fluid: Benzene

Calibrated with 40.0 ppm-m Reference Standard

Calibration Cell Serial # **BENZ052522-06**

Calibration Cell Concentration: **59.6 ppm-m**

Lower 95% confidence: 57.8 ppm-m

Correlation Coefficient: 0.992

Upper 95% confidence: 61.3 ppm-m

Laboratory Manager:

Date: May 25, 2022

Brentley S. Olive MSPH, Ph.D., CIH



DEPARTMENT of CHEMISTRY and INDUSTRIAL HYGIENE

College of Arts and Sciences

UNA Box 5049, Florence, AL 35632-0001

P: 256.765.4622 | F: 256.765.4958 | www.una.edu/chem-ih

Equal Opportunity / Equal Access Institution



Sealed Calibration Cell

Certificate of Calibration

This calibration cell was individually calibrated by spectroscopic analysis using ultra-violet reference libraries created by permeation tubes traceable to N.I.S.T standards.

Permeation Tubes Serial Numbers: 49302/49303

Permeation Fluid: Benzene

Calibrated with 20.0 ppm-m Reference Standard

Calibration Cell Serial # **BENZ052522-04**

Calibration Cell Concentration: **31.2 ppm-m**

Lower 95% confidence: 32.7 ppm-m

Correlation Coefficient: 0.979

Upper 95% confidence: 36.5 ppm-m

Laboratory Manager: _____

Date: May 25, 2022

Brentley S. Olive MSPH, Ph.D., CIH



DEPARTMENT of CHEMISTRY and INDUSTRIAL HYGIENE

College of Arts and Sciences

UNA Box 5049, Florence, AL 35632-0001

P: 256.765.4622 | F: 256.765.4958 | www.una.edu/chem-ih

Equal Opportunity / Equal Access Institution